PrymaLab
PRYMALAB ANALYTICAL · IN-HOUSE LABORATORY
3705 Music Branch Rd, Ashland, KY 41102 · +1 (606) 465-9063 · prymalab.net

CERTIFICATE NO.
PL-26-00030
VERIFICATION KEY
54Z32UB2
Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE

Epitalon 50 mg

Lot **PL-2609-030** · Lyophilized powder · Label claim **50 mg/vial**

TESTED IN-HOUSE BY PRYMALAB
In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-030	2026-09-18	2026-09-19	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-18	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND
Epitalon
Epithalon; Ala-Glu-Asp-Gly (AEDG); Khavinson pineal peptide bioregulator

CAS NUMBER
307297-39-8

MOLECULAR FORMULA
$C_{14}H_{22}N_4O_9$

THEORETICAL MASS (AVG.)
390.4 Da

SEQUENCE
`H-AEDG-OH (4 residues; free N- and C-termini)`

PURITY · RP-HPLC, AREA %
99.35%
Specification 99.0% • main peak at 2.88 min


IDENTITY · OBSERVED MASS
390.6 Da
Theoretical 390.4 Da • +0.2 Da (512 ppm)

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	390.6 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.35%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.27% at 3.62 min (Ala-Glu-Asp tripeptide)	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.03 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: EPI_PL-2609-030_20260919_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; main peak and impurities labeled with retention times. Peak table on page 2.

Verify: prymalab.net/testing/verify · PL-26-00030 · key 54Z32UB2

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · EPI_PL-2609-030_20260919_01

INSTRUMENT
Waters Alliance e2695, SN E2695B742
DATA SYSTEM
Empower 3, Build 3471
DETECTOR
Waters 2998 PDA
DETECTION WAVELENGTH
214 nm
COLUMN
C18, 150 × 4.6 mm, 3.5 µm
COLUMN PART / SERIAL
186003326 / USHX18211
MOBILE PHASE A
0.1% TFA in water
MOBILE PHASE B
0.1% TFA in MeCN
FLOW RATE
1.0 mL/min
COLUMN TEMPERATURE
30 °C
INJECTION VOLUME
20 µL
RUN TIME
10 min
SAMPLE DILUENT
Water, 0.1% TFA
SAMPLE CONCENTRATION
0.5 mg/mL
BLANK INJECTED
Yes
REFERENCE STANDARD
In-house Epitalon reference, lot PL-RS-EPI-01, RT 2.86 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1 (Ala-Glu dipeptide)	2.24	6,020	0.14
2	**2.88**	**4,272,050**	**99.35**
3 (Ala-Glu-Asp tripeptide)	3.62	11,610	0.27
4	4.38	10,320	0.24
Total		4,300,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters Acquity SQD
IONIZATION / MODE
ESI, positive
OBSERVED M/Z (CHARGE)
391.3 (1+ [M+H]) · 413.3 (1+ [M+Na])
DECONVOLUTED MASS
390.6 Da
THEORETICAL MASS
390.4 Da (free-termini tetrapeptide)
DIFFERENCE
+0.2 Da (512 ppm)
DATA FILE
EPI_PL-2609-030_MS_20260919_01
ACCEPTANCE
Within ± 1.0 Da

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-20	RPT-KF-PL-2609-030
Endotoxin	PrymaLab	2026-09-21	RPT-ET-PL-2609-030
Sterility	PrymaLab	2026-10-02	RPT-ST-PL-2609-030
Peptide content	PrymaLab	2026-09-20	RPT-AA-PL-2609-030

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-19

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00030 and key 54Z32UB2 at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.